

Site: _____ Month: _____ **SCHEDULE SYS_D**
 System ID#: _____ Dispatch/Job #: _____
 Field Service Engineer: _____

Procedure	Sec. No.	Tool Serial #	Pass	Cleaned	Adjusted	Replaced	Repaired	Needs Adjustment	Needs Replacement	Needs Repair	Failed
Safety and Regulatory											
Check Cardiac Gating Cable	1-5										
Check Pneumatic Patient Alert System	1-7										
Check RF Output Power	2-6										
Check Emergency Release of Cradle and Patient Transport	3-1										
Check Patient Transport Casters, Armboard Set Screws and Bumper Strips	3-3										
Check Patient Transport Caster Locks	3-4										
Check Cradle Longitudinal Drive Clutch	3-5										
Image Quality											
Check Eddy Current Compensation	4-2										
Check Gradient Calibration: X	4-4										
Y											
Z											
Check Shim	7-3										
Check SNR	7-4										
Other (Signa Horizon 5.X, 8.X, 9.X & 10.X Not Available for Patient Scans)											
Check Patient Blower & Filter	1-6										
Check Erbttec Blower Output	2-2										
Check Patient Transport Docking Alignment	3-2										
Clean Lightweight Cradle Wheels	3-7										
Check PowerTech and/or Transtector	5-4										
Check OC/IC Computer Fans and Clean Air Intake Grills	6A-1										
Check OC/IC Console Power Supply Fan	6A-3										
Check Operator Workspace Fans And Clean Air Intake Grills	6B-1										
Clean Operator Workstation Mouse & VideoCam Lens	6B-2										
Set SGI System Clock	6B-3										

Site: _____ Month: _____ **SCHEDULE SYS_D**
 System ID#: _____ Dispatch/Job #: _____
 Field Service Engineer: _____

Procedure	Sec. No.	Tool Serial #	Pass	Cleaned	Adjusted	Replaced	Repaired	Needs Adjustment	Needs Replacement	Needs Repair	Failed
Check and Delete Error/Message Log/T-Test Files & Check Date of SaveInfo	7-2										
Perform Site Restoration - Check DQA Phantom, Check Alignment Lights, Remove GE Test Scans, and Check Cabinet Doors and Covers	7-9										
Check Modem	7-10										
Check Laser Camera Fans	8-1										
Clean Laser Camera Suction Cups	8-3										
Clean Laser Camera Transport Plate/Docking Unit	8-4										
Run Laser Camera Cleaning Film	8-5										
Replace Laser Camera Docking Switches	8-8										
Other (Signa Horizon5.X, 8.X, 9.X & 10.X Available for Patient Scans)											
Check RF Cabinet Fans and Clean Filters	2-1										
Check Gradient Cabinet Fans and Filters	4-1										
Check Fluid Level and Valve of Heat Exchanger for Gradient Coil Cooling	4-5										
Check Water Chiller for Gradient Coil Cooling	4-6										
Check Water Chiller Pump Motor Lubrication	4-7										
Check Standard PDU Fans and Filters	5-1										
Check Compact PDU Fans and Filters	5-5										
Clean OC/IC DAT Drive	6A-2										
Check System Cabinet Fans and Clean Filters	7-1										
Check Cabinet Inlet Air Temperature	7-5										
Verify PM Completion on Van Equipment (Mobiles Only)	7-8										
Review System Health Report	7-11										
TAC Filter Replacement	9-1										

Notes: _____

GE HEALTHCARE

Site: _____ Month: _____ **SCHEDULE SYS_D**
System ID#: _____ Dispatch/Job #: _____
Field Service Engineer: _____

FSE ID #: _____

Field Service Engineer: _____ Date: _____