SUBSYSTEM INSTALLATION

Locate the each sub-system to the proper location.



Note

It is not necessary to locate the Table in front of the magnet before 7th day ”Table Installation”.

Locate the Table so that it does not disturbe the installation procedure